HELLENIC ACCREDITATION SYSTEM – ESYD

ESYD STATEMENT REGARDING THE
REISSUE OF CERTIFICATES AND REPORTS (ISO/IEC 17025)

According to ISO/IEC 17025, clause 7.8.8.1, where an amendment, modification, or reissue of a calibration certificate or test report is required, the additional information shall be clearly identified and, where applicable, the reason for such modification shall be stated in the report/certificate.

It is noted that calibration certificates / test reports shall be reissued only for corrections of errors, as well as for the addition of data / information that had been omitted but were available at the time the calibration / test was performed.

Furthermore, any amendment, modification, or reissue of certificates / reports requested by the customer shall not affect the integrity of the results, and any modifications to the reported results, for whatever reason, shall relate to the item that was tested / calibrated. For example, it is not acceptable to add / change data if they do not relate to the item / test specimen / equipment referred to in the initial report.

Dr. I. Sitaras
Director of Laboratories Accreditation Directorate, ESYD.

Aik. Bouranta
Head of Calibration Laboratories and Other Metrological Services Department

M. Papatzikou
Quality Director, ESYD