Hellenic Accreditation System

Annex F2/5 to the Certificate No. **835**-4

SCOPE of ACCREDITATION

of the

Calibration Laboratory

of

LOGICOM S.A.

Calibration item / Parameters	Range of measurement	Expanded measurement uncertainty (k=2) *	Remarks
Fixed metallic fuel tanks	0,5 m^3 to 200 m^3	0,06 %	Volumetric method using volumetric vessels according to guideline EURAMET cg-21, v3.0: 2024 Calibration medium: tap water Calibration is performed on site.

* *Where the expanded uncertainty (with coverage probability 95 %) is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative.*

The Calibration & Measurement Capability (CMC) includes the measurand, the measurement range and the measurement uncertainty and expresses the smallest measurement uncertainty that can be achieved in a calibration.

Permanent laboratory premises address: **109, Marathonos Ave., 153 44 Gerakas, Greece**
Approved Signatories: **N. Tsakiris**

This Scope of Accreditation replaces the previous one dated 09.09.2024.
The Accreditation Certificate No.**835**-4, to ELOT EN ISO/IEC 17025: 2017, is valid until 27.09.2029.

Athens, 21.03.2025

Konstantinou Evangelos Apostolos
CEO of ESYD