Hellenic Accreditation System

Annex F1/9 to the Certificate No. **196**-8

SCOPE of ACCREDITATION

of the

Testing and Calibration Laboratory

of

'CQS S.A. – Calibration and Quality Services'

Materials / Products to be tested	Type of Tests / Properties to be measured	Applied Methods / Techniques to be used
Spectrophotometer performance tests		
Ultraviolet & Visible Spectrophotometers **(**UV – Vis)	Wavelength accuracy & precision in the range of 235 nm to 875 nm	ASTM E 275-08(2022) The testing is also performed on site
	Stray Light at wavelengths 261 nm and 350 nm	ASTM E 387-04(2022) The testing is also performed on site
	Linearity of detector response in the range 235 nm to 350 nm	ASTM E 275-08(2022) The testing is also performed on site
	Resolution (A269/A266)	European pharmacopoeia 12.0, 2.2.25, Absorption Spectrophotometry Ultraviolet and Visible The testing is also performed on site

Site of assessment: **Permanent laboratory premises, 44 Ag. Triados Str., Pefka, 57010 Thessaloniki, Greece.**
Approved Signatories: **A. Kougioumtzidis, C. Vratmiras**

This Scope of Accreditation replaces the previous one dated September 21st, 2025.
The Accreditation Certificate No. **196**-8, to ELOT EN ISO/IEC 17025:2017, is valid until 22.03.2030.

Athens, October 10th, 2025

Konstantinou Evangelos Apostolos
CEO of ESYD