Hellenic Accreditation System

Annex F2/8 to the Certificate No. **1002**-4

SCOPE of ACCREDITATION

of the

Calibration Laboratory

of

SKOPELITIS LEMONIS

Calibration item / Parameters	Range of measurement	Expanded Measurement Uncertainty (k=2)*	Remarks
Volume measurements			
Fixed metallic tanks	1 m^3 to 100 m^3	0,1 % of nominal tank volume	Volumetric method Guideline followed: EURAMET cg-21 Version 2.1 (09/2021) Calibration medium: Tap water The calibration can be performed on site.
Road metallic tanks	0,5 m^3 to 40 m^3	0,08 % of nominal tank volume	Volumetric method Calibration medium: Tap water Guidelines followed: EURAMET cg-21 Version 2.1 (09/2021) OIML R80-2: 2017, Annex B The calibration can also be performed on site.

** Where the expanded uncertainty (with 95 % coverage) is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative.*

The Calibration Measurement Capability (CMC) includes the measured quantity, the measurement range and the measurement uncertainty, expressing the minimum measurement uncertainty which can be achieved in a calibration.

Permanent laboratory premises address: **4th klm of National Road Volos-Larisa, 38 500, Volos, Greece.**

Approved Signatories: **L. Skopelitis, N. Skopelitis**

This Scope of Accreditation replaces the previous one dated 17.06.2024.
The Accreditation Certificate No. **1002**-4, to ELOT EN ISO/IEC 17025:2017, is valid until 04.10.2028.

Athens, 02.09.2024

Christos Nestoras
CEO of ESYD