Hellenic Accreditation System

Annex F2/8 to the Certificate No. **995**-3

SCOPE of ACCREDITATION

of the

Calibration Laboratory

XANTHOPOULOS PANTELIS

Calibration item/ Parameters	Range of measurement	Expanded measurement Uncertainty-(k=2)*	Remarks
Volume measurements			
Fixed Metallic Tanks	[0,5 to 100] m^3	0.1 % of nominal tank volume	According to standard & guide: ISO 4269:2001 EURAMET cg-21 Version 2.1 (09/2021) Calibration medium: Tap water. The calibration is performed on site.

**Where expanded uncertainty (with a probability of 95% coverage) is accompanied by the corresponding unit is absolute, while where it is not accompanied by a unit, it is relative.*
The Calibration Measurement Capability (CMC) includes the measured quantity, the measurement range and the measurement uncertainty, and expresses the smallest measurement uncertainty which can be achieved in a calibration.

Permanent laboratory premises address**: Nea Kessani, 67 063, Xanthi, Greece.**
Approved Signatory: **P. Xanthopoulos**

This scope of Accreditation replaces the previous one dated 29.01.2024.
The Accreditation Certificate No. **995**-3, to ELOT EN ISO/IEC 17025:2017, is valid until 03.08.2028.

Athens, 09.04.2024

Christos Nestoras
CEO of ESYD