HELLENIC ACCREDITATION SYSTEM

Annex G1/7 to the Certificate No. **1038**-4

SCOPE of ACCREDITATION

of the

Clinical Laboratory

of

AFFIDEA Private Diagnostic Laboratories and Multimodalities Medical Single MEMBER S.A. (AFFIDEA of DAFNI)

Materials / Products tested	Types of test / Properties measured	Applied methods / Techniques used
Hematological Tests		
Whole blood (EDTA)	**Determination of 8 parameters**	**Automated analyzer SYSMEX XN-530** **OE 504-11_3/20.5.2026**
	1. White Blood Cells (WBC)	Flow Cytometry with fluorescence method
	2. Red Blood Cells (RBC)	Hydrodynamic Focusing (DC Detection)
	3. Hemoglobin (HGB)	Photometry
	4. Hematocrit (HCT)	Hydrodynamic Focusing (DC Detection)
	5. Mean Cell Volume (MCV)	Calculated based on specific equation (DC Detection)
	6. Mean Cell Hemoglobin (MCH)	Calculated based on specific equation (DC Detection)
	7. Mean corpuscular hemoglobin concentration (MCHC)	Calculated based on specific equation (DC Detection)
	8. Platelets (PLTs)	Hydrodynamic Focusing (DC Detection)

**Reference to the commercial name of a specific analyzer/kit, refers to a specific analytical method and protocol*

Primary Sample Collection Site: **226 Vouliagmenis Avenue, Ag. Dimitrios, Attiki, Greece**

Approved signatories : **Theodora Panouklia.**

This scope of Accreditation replaces the previous one dated 18.07.2025.
The Accreditation Certificate No. **1038-**4, to ELOT EN ISO 15189:2022, is valid until **08.05.2029**.

Athens, 31st August 2026

Konstantinou Evangelos Apostolos
CEO of ESYD