Hellenic Accreditation System

Annex F1/13 to the Certificate No. **388**-5

SCOPE of ACCREDITATION

of the

Laboratory

of

CHIMIKI TECHNIKI MINESCHOS L.P.

for the performance of tests

Materials/ Products to be tested	Types of test/ Properties to be measured	Applied methods/ Techniques to be used
Performance tests		
Visible and ultra-violet spectrophotometers (UV-Vis)	Wavelength accuracy (200 nm - 900 nm)	ASTM E 275: 2008 (R2022) Test is also performed on-site.
	Stray light (220 nm)	ASTM E 387: 2004 (R2022) Test is also performed on-site.
	Resolution (A268,7/A267,0)	NIST Special Publication 260-54 EurPharm 2.2.25 (11th ed./2024) Test is also performed on-site.
Visible and ultra-violet detectors (UV-Vis) of high-performance liquid chromatographs (HPLC)	Wavelength accuracy (240 nm - 650 nm) Linear range	ASTM E 1657: 1998 (2019) Test is also performed on-site.
Conductivity detectors of ion chromatographs (IC) and high-performance liquid chromatographs (HPLC)	Linear range	ASTM E 1511: 1993 (2017) The testing is also performed on site.

Site of assessment: **Permanent laboratory premises, 197, Spaton Avenue, GR-153 51 Pallini, Attiki, Greece.**
Approved signatories: **Gerasimos Mineschos.**

This Scope of Accreditation replaces the previous one dated October 4th, 2023.
The Accreditation Certificate No. **388**-5, to ELOT EN ISO/IEC 17025:2017, is valid until 09.10.2028.

Athens, March 15th, 2024

Christos Nestoras
CEO of E.SY.D