Hellenic Accreditation System

Annex F1/9 to the Certificate No. **427**-5

SCOPE of ACCREDITATION

of the

Water Quality Control Laboratory

of the

Water Sewage Municipal Agency of Greater Volos (DEYAMV)

Materials/ Products tested	Types of test/ Properties measured	Applied standards/ Techniques used
Chemical Tests		
Water	1. Determination of electrical conductivity	ELOT EN 27888
	2. Determination of pH	ELOT 658 :1983
	3. Determination of chlorides	ELOT 666 :1983 modified method
Microbiological Tests		
Potable and Swimming Pool Water	1. Detection and enumeration of *Escherihia coli* and coliform bacteria	ISO 9308-1:2014
	2. Detection and enumeration of intestinal enterococci	ISO 7899-2:2000
	3. Enumeration of culturable micro-organisms	ISO 6222:1999

Site of assessment: **Permanent laboratory premises – DEYAMV (Hatziargiri), Kessarianis and Alexandrou Merou st, 38333, Volos, Greece**

Approved Signatories: **Mavrokefalidou Niki, Dakis Dimitrios**

This scope of Accreditation replaces the previous one dated 11.11.2024.
The Accreditation Certificate No **427**-5, to ELOT EN ISO/IEC 17025:2017, has been renewed until 14.05.2028.

Athens, 13th of May 2025

Konstantinou Evangelos Apostolos
CEO of ESYD