Hellenic Accreditation System

Annex G1/8 to the Certificate No.**948**-6

SCOPE of ACCREDITATION

of the

Clinical Laboratory

of

AFFIDEA OF VARI

Private Multimodality, Medical Single Member S. A.

Materials/Products tested	Types of test/Properties measured	Applied methods/Techniques used
Hematological Tests		
Whole blood	**Determination of 8 parameters**	**Automated analyzer SYSMEX XN-L 550*** **OE 504-11_2/14.7.2016**
	1. White Blood Cells (WBC)	Flow Cytometry with fluorescence method
	2. Red Blood Cells (RBC)	Hydrodynamic Focusing (DC Detection)
	3. Hemoglobin (HGB)	Photometry
	4. Hematocrit (HCT)	Hydrodynamic Focusing (DC Detection)
	5. Mean Cell Volume (MCV)	Calculated based on specific equation (DC Detection)
	6. Mean Cell Hemoglobin (MCH)	Calculated based on specific equation (DC Detection)
	7. Mean corpuscular hemoglobin concentration (MCHC)	Calculated based on specific equation (DC Detection)
	8. Platelets (PLTs)	Hydrodynamic Focusing (DC Detection)

**Reference to the commercial name of a specific analyzer/kit, refers to a specific analytical method and protocol*

Site of assessment : **Laboratory permanent premises, 40-42 Varis-Koropiou Av., Vari, Attiki, Greece.**
Approved signatories : **Fedra Kontaxaki, Eleni Douganioti.**

This Scope of Accreditation replaces the previous one dated 12.04.2021.
The Accreditation Certificate No.**948**-6, to ELOT EN ISO 15189:2012, is valid until 17.11.2027.

Athens, 30.11.2022

Christos Nestoras
CEO of ESYD