Hellenic Accreditation System

Annex F2/4 to the Certificate No. **1258**

SCOPE of ACCREDITATION

of the

Calibration Laboratory

of

KAISIS OIL GROUP S.A. (NICK TRANS S.A.)

Calibration item / Parameters	Range of measurement	Expanded Uncertainty of Measurement(k=2)*	Remarks
Volume measurements			
Fixed metallic tanks	1 m^3 to 200 m^3 (maximum tank height 3m)	0,12 % of nominal tank volume	Calibration according to : EURAMET cg-21 Version 2.1 (09/2021) Calibration medium: Tap water The calibration can be performed on site.
Road metallic tanks	0,5 m^3 to 40 m^3	0,12 % of nominal tank volume	Volumetric method through the use of volumetric standard Calibration medium: Tap water Standards followed: EURAMET cg-21 Version 2.1 (09/2021) OIML R80-2: 2017, Annex B The calibration can also be performed on site

** Where the expanded uncertainty (with 95 % coverage) is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative.*

The Calibration Measurement Capability (CMC) includes the measured quantity, the measurement range and the measurement uncertainty, expressing the minimum measurement uncertainty which can be achieved in a calibration.

Permanent laboratory premises address: **Proektasi Makrigianni, Evosmos, Thessaloniki, Greece.**
Approved Signatories: **Dimitrios Polychronakos, Christos Zikidis.**

This Scope of Accreditation replaces the previous one dated March 12th, 2025.
The Accreditation Certificate No. **1258**, to ELOT EN ISO/IEC 17025:2017, is valid until 07.10.2025, according to ESYD decision.

Athens, March 17th, 2025

Konstantinou Evangelos Apostolos
CEO of ESYD