Hellenic Accreditation System

Annex F1/9 to the Certificate No.**1139-**5

SCOPE of ACCREDITATION

of the

Sewage Quality Control Laboratory

of the

Analytical Laboratories Division

of the

Thessaloniki Water Supply & Sewerage Company S.A. (EYATH SA)

Tested materials/products	Types of tests/ Properties to be measured	Applied Standards/Techniques to be used
Chemical Tests		
Industrial Liquid Wastes & Wastewaters	1.Determination of pH	In house method based on ELOT 658/1983 *Laboratory method code: MA-01*
	2.Determination of Total Suspended Solids (TSS)	In house method based on 2540 D APHA (24rd Ed., 2023) *Laboratory method code: MA-02*
	3.Determination of Chemical Oxygen Demand (COD)	Merck 1.14540.0001 and 1.14541.000 Laboratory method code: MA-03
	4.Determination of NH_4-N (Ammonium Nitrogen)	Merck 1.14558.0001 and 1.14559.0001 Laboratory method code: MA-04
	5.Determination of NO_3-N (Nitrate Nitrogen)	Merck 1.14563.0001 Laboratory method code: MA-05
	6.Determination of Total Nitrogen (TN)	Merck 1.00613.0001 and 1.14763.0001 Laboratory method code: MA-06
	7.Determination of Total Phosphorous (TP)	Merck 1.14729.0001 *Laboratory method code: MA-07*

Tested materials/products	Types of tests/ Properties to be measured	Applied Standards/Techniques to be used
	8.Determination of True Color (TC)	In house method based on 2120 C APHA (24rd Ed., 2023) *Laboratory method code: MA-09*
Non-disinfected Industrial Liquid Wastes & Wastewaters	Determination of Biological Oxygen Demand (BOD_5)	In house method based on 5210 B APHA (24rd Ed., 2023) *Laboratory method code: MA-08*

Site of assessment: **Sewerage Quality Control Laboratory, 98 Tsimiski Street, Thessaloniki GR-54623.**
Approved signatories: **G. Nikolaidis, N. Filippidis, A. Sakellariou.**

This scope of Accreditation replaces the previous one dated 14.01.2026.
The Accreditation Certificate No.**1139-**5, according to ELOT ISO/IEC 17025: 2017, is valid until 12.09.2027.

Athens, 07.05.2026

Konstantinou Evangelos Apostolos
CEO of ESYD