\Hellenic Accreditation System

Annex F2/2 to the Certificate No. **1334**

SCOPE of ACCREDITATION

of the

Calibration Laboratory

of

QAFA GRAMOZ - SOLUTIONS PLIROFORIKI

Calibration item / Parameters	Range of measurement	Expanded measurement uncertainty (k=2)*	Remarks
Volume measurements			
Steel tanks containing: - Fuel - Alcoholic bevarages	0,2 m^3 to 50 m^3	0,1 % (filling tank) 0,11 % (emptying tank) of nominal tank volume	Volumetric method using volume standard according to: EURAMET cg-21, v.3, 2024 Calibration medium: potable water. Calibration is performed on site.
Road tankers	0,2 m^3 to 20 m^3	0,1 % of nominal tank volume	Volumetric method using volume standard according to: - EURAMET cg-21, v.3, 2024 - OIML R80-2:2017, Annex B Calibration medium: potable water. Calibration is performed at laboratory premises or on site.

** Where expanded uncertainty (with a probability of 95% coverage) is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative.*

The Calibration & Measurement Capability (CMC), includes the measured quantity, measurement range and measurement uncertainty and expresses the smallest measurement uncertainty that can be achieved during a calibration.

Permanent laboratory premises address: **Likomidous 5, 26333 Patra.**
Approved Signatories: **G. Qafa, T. Nikolaou**

This Scope of Accreditation replaces the previous one dated March 15th, 2023
The Accreditation Certificate No. **1334**, to ELOT EN ISO/IEC 17025:2017, is valid until March 14th, 2028.

Athens, June 3rd, 2025

Konstantinou Evangelos Apostolos
CEO of ESYD