Hellenic Accreditation System

Annex F1/8 to the Certificate No. **789**-5

SCOPE of ACCREDITATION

of the

Calibration & Testing Laboratory

of

N. Asteriadis S.A.

for the performance of tests

Materials / Products to be tested	Type of test / Properties to be measured	Applied methods / Techniques to be used
Performance Testing		
UV-Vis Spectrophotometers	Wavelength accuracy (240 nm to 642 nm)	ASTM E 275:2008 ASTM E 925:2009 Testing is also performed on-site.
	Wavelength repeatability (240 nm to 642 nm)	ASTM E 275:2008 Testing is also performed on-site.
	Stray light (190nm to 220 nm)	European Pharmacopoeia Edition 11.0, July 2022, 2.2.25. Ultraviolet and Visible Absorption Spectrophotometry Testing is also performed on-site.
	Resolution (A269/A266)	

Site of assessment: **Permanent laboratory premises, 31 Dervenion Str. & Poseidonos Str., GR-14451, Metamorfossi, Attiki, Greece**

Approved Signatories: **Emmanouil Barbounis, Konstantinos Tampouris, Fotios Fotiadis.**

This Scope of Accreditation replaces the previous one dated March 15th, 2024.

The Accreditation Certificate No. **789**-5, to ELOT EN ISO/IEC 17025:2017, is valid until 23.07.2029.

Athens, August 2nd, 2024

Christos Nestoras
CEO of ESYD