Hellenic Accreditation System

Annex F1/11 to the Certificate No. **887**-4

SCOPE of ACCREDITATION

of the

Testing & Calibration Laboratory

of

METRISIS CALIBRATIONS P.C.

for the performance of tests

Materials / Products to be tested	Types of test / Properties to be measured	Applied methods / Techniques to be used
Performance tests of UV-Vis spectrophotometers		
Visible and ultra-violet spectrophotometers (UV-Vis) in the range 200 nm to 900nm	Baseline stability 200 nm to 900 nm	ASTM E 275-08(2022) Testing is also performed on-site.
	Stray light 279 nm, 465 nm	European Pharmacopoeia Edition 10.0 Jan 2020, 2.2.25 Testing is also performed on-site.
	Resolution A268,8/A266,8	
	Wavelength accuracy (Holmium filter) 230 nm to 640 nm Wavelength accuracy (Didymium filter) 330 nm to 880 nm	

Site of assessment: **Permanent premises, 95 Lambrou Katsoni str, 183 44, Moschato, Piraeus.**
Approved Signatories: **C. Konstantopoulos, G. Fardis.**

This Scope of Accreditation replaces the previous one dated 12.03.2026.
The Accreditation Certificate No. **887**-4, to ELOT EN ISO/IEC 17025:2017, has been extended according to ESYD's decision and it is valid until 29.07.2026.

Athens, 21.05.2026

Konstantinou Evangelos Apostolos
CEO of ESYD