HELLENIC ACCREDITATION SYSTEM

Annex F2/13 to the Certificate No. **922**-4

SCOPE of ACCREDITATION

of the

Calibration Laboratory

of

DOMIKI PROODOS S.A.

Measurand / Calibration item	Range of measurement	Calibration & Measurement Capability (k=2)*	Remarks
Volume measurements			
Fixed storage tanks	0,5 m^3 to 50 m^3	0,06 % of nominal tank volume	Volumetric method using volumetric vessel according to: EURAMET cg-21, v.2.1: 2021 Calibration medium: Tap water. Calibration is performed on site.
Fixed storage tanks, including tanks of alcoholic beverages	0,2 m^3 to 750 m^3	0,08 % of nominal tank volume	Volumetric method using flow meter according to: ISO 4269:2001 Calibration medium: Tap water or other liquid with similar volatility and viscosity or suitable petroleum product. Calibration is performed on site.
Road tankers	0,2 m^3 to 80 m^3	0,08 % of nominal tank volume	Volumetric method using flow meter according to: - ISO 4269:2001 - OIML R 80-2: 2017 - Annex B Calibration medium: Tap water or other suitable petroleum product. Calibration is performed on site.
Transportable tanks	0,2 m^3 to 80 m^3	0,08 % of nominal tank volume at maximum capacity of the tank	Volumetric method using flow meter according to: - ISO 4269:2001 - OIML R 80-2: 2017 - Annex B Calibration medium: Tap water or other suitable petroleum product. Calibration is performed on site.

** Where uncertainty is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, is relative*

Site of assessment: **Permanent laboratory premises, Petras str 3, 12241, Aigaleo, Attiki, Greece.**

Approved Signatory: **Nikolaos Ventouris, Christos Boubousis.**

This Scope of Accreditation replaces the previous one dated 23.04.2026.
The Accreditation Certificate No. **922**-4, to ELOT EN ISO/IEC 17025: 2017, is valid until 14.09.2027.

Athens, 15.07.2026

Konstantinou Evangelos Apostolos
CEO of ESYD