Hellenic Accreditation System

Annex F2/9 to the Certificate No. **974**-4

SCOPE of ACCREDITATION

of the

Tank Calibration Laboratory

of

Athanasios Mylonas

Calibration item / Parameters	Range of measurement	Expanded Measurement Uncertainty (k=2)*	Remarks
Liquid volume / Fixed metallic tanks	0.5 m^3 to 210 m^3	0.1 % of nominal tank volume	Volumetric method through the use of a flow meter Calibration medium: Tap water Standard followed: ISO 4269: 2001 The calibration can be performed on site
Liquid volume / Road metallic tanks	0,5 m^3 to 40 m^3	0.15 % of nominal tank volume	Volumetric method through the use of a flow meter Calibration medium: Tap water Standards followed: ISO 4269: 2001 OIML R80-2: 2017, Annex B The calibration can also be performed on site

** Where the expanded uncertainty (with 95 % coverage) is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative.*

The Calibration Measurement Capability (CMC) includes the measured quantity, the measurement range and the measurement uncertainty, expressing the minimum measurement uncertainty which can be achieved in a calibration.

Permanent laboratory premises address: **Agamemnonos str. 6, Dentropotamos, Thessaloniki.**
Approved Signatories: **A. Mylonas.**

This Scope of Accreditation replaces the previous one dated 12.07.2023.

The Accreditation Certificate No. **974**-4, to ELOT EN ISO/IEC 17025:2017, is valid until 21.04.2028.

Athens, 18.12.2023

Christos Nestoras
CEO of ESYD