Hellenic Accreditation System

Annex G1/1 to the Certificate No. **1501**

SCOPE of ACCREDITATION

of the

Clinical laboratory

of

Private Polyclinic in eastern Thessaloniki, Evrodiagnosi

Tested materials / products	Types of test / Properties	Applied standards / Techniques
Biochemical tests		
Blood serum	**Determination of 15 parameters**	**Automated biochemical analyzer, Alinity c-series, Abbott (E.703A-2/1st/11.03.2026)**
	1. Glucose	Enzymatic Colorymetric(G-6-PDH)
	2. Urea	Enzymatic Colorymetric (Urease)
	3. Uric acid (UA)	Enzymatic Colorymetric (Urease)
	4. Aspartate aminotransferase (AST/GOT)	Enzymatic kinetic colorimetric (NADH)
	5. Alanine aminotransferase (ALT/SGPT)	Enzymatic kinetic colorimetric (NADH)
	6. Gamma glutamyl transferase (γGT)	Enzymatic Colorymetric
	7. Total cholesterol (tCHOL)	Enzymatic Colorymetric
	8. High Density Lipoprotein Cholesterol (HDL-Chol)	Enzymatic Colorymetric (ASD)
	9. Triglycerides (TRIG)	Enzymatic Colorymetric (GPO)
	10. Sodium (Na)	Potentiometric (ISE)
	11. Potassium (K)	Potentiometric (ISE)
	12. Iron	Colorimetric
	13. Creatinine kinase (CK)	Enzymatic (NAC)
	14. Albumin (ALB)	Colorimetric (BCG)
	15. Total Protein (TP)	Colorimetric (Biuret)

**Reference to the commercial name of a specific analyzer, refers to a specific analytical method and protocol*
***IFCC: International Federation of Clinical Chemistry*

Site of assessment: **Permanent laboratory premises, Adrianoupoleos 1, Kalamaria, Thessaloniki.**
Sites of primary sample collection: **1. Adrianoupoleos 1, Kalamaria, Thessaloniki (central laboratory)**
2. Metamorfoseos 26, Kalamaria, Thessaloniki (branch)
3. Pavlou Mela 5 (Evosmos pedestrian street), Evosmos, Thessaloniki (branch)
Approved signatories: **Ioanna Tzima, Aristeidis Zaires, Vasileios Katsirmas, Marianthi Pagouni**

The Accreditation Certificate No. **1501**, to ELOT EN ISO 15189:2022, is valid until 27.05.2031.

Athens, 28th May 2026

Konstantinou Evangelos Apostolos
CEO of ESYD